



SCALE TEST REPORT

GTR SCALES LTD.
 196 Fleming Drive
 Burnstown Ontario
 K0J 1G0

CUSTOMER: EAA Chapter 245

ORDER# 210728EAA245

MAKE	MODEL	SERIAL #	LOCATION	CAPACITY	TEST LOAD APPLIED				MAX. ERROR 4th LOAD	PASS FAIL	CORRECTION MADE YES/NO
					1st LOAD	2nd LOAD	3rd LOAD	4th LOAD			
1	Weigh Systems Inc	Pad Scale	882	Before	2000lb x 0.5lb	1000lb	2000lb				
						1001.0	2001.5				
2	Weigh Systems Inc	Pad Scale	882	After	2000lb x 0.5lb	1000lb	2000lb				
						1000.5	2001.0				
3	Weigh Systems Inc	Pad Scale	883	Before	2000lb x 0.5lb	1000lb	2000lb				
						1002.0	2004.0				
4	Weigh Systems Inc	Pad Scale	883	After	2000lb x 0.5lb	1000lb	2000lb				
						1000.0	2000.0				
5	Weigh Systems Inc	Pad Scale	884	Before	2000lb x 0.5lb	1000lb	2000lb				
						1002.0	1002.5				
6	Weigh Systems Inc	Pad Scale	884	After	2000lb x 0.5lb	1000lb	2000lb				
						1000.0	2000.5				
7											
8											
9											
10											

WEIGHT ID #Set B

CERTIFICATE # 1352400

COMMENTS:

Technician Name: James Kucharik

Technician Signature: *JK*

CALIBRATION DONE: September 28 2021

Customer Name: Phillip Johnson

Customer Signature:

CALIBRATION DUE: September 2022